

LOT NO. 8121912-9

EXP:97/10/01

INSTRUMENT: SYSMEX K₁₀₀₀ & K₈₀₀ & K₄₅₀₀

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MXD	NEUT
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0,9	±0,20	±0,6	±2,4	±6	±2,3	±2	±0	±20	±2,3	±0	±2	±0
MEAN	7,6	4,48	13,2	36	81,0	29,0	36,0	13,0	248	11	37	7	06

INSTRUMENT: SYSMEX KX₂₁ & N/XP₁₀₀ & 300

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MXD	NEUT
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0,9	±0,20	±0,6	±2,4	±6	±2,3	±2	±0	±20	±2,3	±0	±2	±0
MEAN	7,6	4,49	13,3	36	81,7	29,0	36,0	13,0	200	11	37	7	06

INSTRUMENT: DIATRON

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0,9	±0,20	±0,6	±2,4	±6	±2,3	±2	±0	±20	±2,3	±0	±0	±0
MEAN	7,0	4,40	13,1	30,0	81,0	29,0	36,0	13,0	240	11	37	3	60

INSTRUMENT: ORPHEE 2D

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0,9	±0,20	±0,6	±2,4	±6	±2,3	±2	±0	±20	±2,3	±0	±0	±0
MEAN	7,6	4,47	13,2	30,6	81,0	29,0	36,0	13,0	246	11	37	3	60

INSTRUMENT: MICROS/ABX

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0,9	±0,20	±0,6	±2,4	±6	±2,3	±2	±0	±20	±2,3	±0	±0	±0
MEAN	7,0	4,44	13,1	30,4	80,0	29,1	36,0	13,0	230	11	37	3	60

INSTRUMENT: MINDRAY 2800-2000

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0,9	±0,20	±0,6	±2,4	±6	±2,3	±2	±0	±20	±2,3	±0	±0	±0
MEAN	7,6	4,49	13,3	36	81,7	29,0	36,0	13,0	200	11	37	3	60

INSTRUMENT: MINDRAY 0800

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl
LIMIT	±0,9	±0,20	±0,6	±2,4	±6	±2,3	±2	±0	±20	±2,3
MEAN	8,30	4,46	12,9	36,8	84	29	30	12,8	230	12,6

LOT NO . 8121912-9

EXP:97/10/01

INSTRUMENT: CELLDYN--

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0.9	±0.20	±0.6	±2.4	±6	±2.2	±2	±0	±20	±2.2	±0	±0	±0
MEAN	7.0	4.44	12.1	30.0	81.0	29.0	36.0	12.0	242	11	27	2	6.0

INSTRUMENT: NIHON KOHDEN

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0.9	±0.20	±0.6	±2.4	±6	±2.2	±2	±0	±20	±2.2	±0	±0	±0
MEAN	7.0	4.40	12.1	30.0	81.0	29.0	36.0	12.0	242	11	27	2	6.0

INSTRUMENT: ERMA -- CONVERGENT

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0.9	±0.20	±0.6	±2.4	±6	±2.2	±2	±0	±20	±2.2	±0	±0	±0
MEAN	7.0	4.46	12.2	30.6	80.0	29.0	36.0	12.0	238	11	27	2	6.0

INSTRUMENT: MEDONIC

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV	LYMPH	MON	GRA
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl	%	%	%
LIMIT	±0.9	±0.20	±0.6	±2.4	±6	±2.2	±2	±0	±20	±2.2	±0	±0	±0
MEAN	7.0	4.40	12.1	30.4	81.0	29.0	36.0	12.0	244	11	27	2	6.0

INSTRUMENT: ORPHEE 0D

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L
LIMIT	±0.9	±0.20	±0.6	±2.4	±6	±2.2	±2	±0	±20
MEAN	4.9	4.40	12.2	30.0	81	29.0	36.4	12	240

INSTRUMENT: Sysmex XT

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	PLT
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	10 ⁹ /L
LIMIT	±0.9	±0.20	±0.6	±2.4	±6	±2.2	±2	±20
MEAN	7.7	4.40	12.2	30.7	81.7	29.0	36.0	240

INSTRUMENT: Mindray 0D

PARAMETER	WBC	RBC	HGB	HCT	MCV	MCH	MCHC	RDW/CV	PLT	MPV
UNIT	10 ⁹ /L	10 ¹² /L	g/dl	%	fl	pg	g/dl	%	10 ⁹ /L	fl
LIMIT	±0.9	±0.20	±0.6	±2.4	±6	±2.2	±2	±0	±20	±2.2
MEAN	7.6	4.46	12.9	30.8	80.0	29	36	12.8	220	12.6