

11. Assay Procedure Continued...

- After 30 minutes, discard the well contents and carefully wash the wells 4 times with Wash Buffer. Ensure that the wells are empty but do not allow to dry out.
- Using a repeating dispenser, rapidly dispense 100µl of TMB Substrate (**Reagent 4**) into each well. Incubate the plate for **10 minutes**.
- Add 100µl of Stop Solution (**Reagent 5**) to each well. To allow equal reaction times, the Stop Solution should be added to the wells in the same order as the TMB Substrate.
- Read the optical density (OD) of each well at 450nm in a microplate reader within 10 minutes. A 620nm filter may be used as a reference wavelength.

12. Quality Control

Quality control data is supplied on the lot-specific QC certificate included in the kit.

Controls are intended to monitor for substantial reagent failure.

Any well positive by spectrophotometer but without visible colour should be cleaned on the underside and re-read. If OD values below zero are observed, the wavelengths used should be verified, the reader re-blanked to air and the measurements repeated.

13. Interpretation of Results

Semi-Quantitative Results

Plot the OD of each standard against its concentration and draw a line through the points. Read the unknowns off this line. Values below 10 U/ml are considered normal. Values between 10 - 12 U/ml are indeterminate. Values above 12 U/ml are regarded as positive.

Qualitative Results

ODs below that of the 10 U/ml standard are considered normal. Samples giving ODs greater than the 10 U/ml standard are regarded as positive.

14. Limitations of the Procedure

- For diagnostic purposes, *H.pylori* IgM ELISA results should be used in conjunction with other test results and overall clinical presentation.
- The sample diluent has been formulated to resolve specific IgG and rheumatoid factor interference. However interference by these antibodies cannot be excluded in samples with very high levels.

15. Performance Characteristics

Under evaluation.

16. Reproducibility

Within Assay Precision

CV%: <10%

Between Assay Precision

CV%: <12%

17. Method Summary

- Mix IgG absorbent and Sample Diluent (**Reagent 1**) 1:4 and dilute all samples 1:100
- Dispense Standards as required, the Positive and Negative Controls and the diluted sample into the microplate wells
- Incubate for **30 minutes** at room temperature.
- Wash the wells three times*
- Dispense 100µl of Conjugate (**Reagent 3**) into each well
- Incubate at room temperature for **30 minutes**
- Wash the wells four times*
- Add 100µl of TMB Substrate (**Reagent 4**) to each well
- Incubate at room temperature for **10 minutes**
- Add 100µl Stop Solution (**Reagent 5**) to each well
- Read the optical density at 450nm (single wavelength) or 450/620nm (dual wavelength).

18. Further Reading

Warren J R & Marshall B J, *Lancet* 1983, I, 1273-1275
Goodwin C S *et al*, *Int. J. Syst. Bacteriol.* 1989, 39, 397 - 405
Rathbone B J *et al*, *Gut* 1986, 27, 635-641
Wyatt J L *et al*, *Scand. J. Gastroenterology* 1989, 24 (Sup 160), 27-34
von Wuffen H & Grote H J, *Euro. J. Clinical. Microbiol. Infect. Dis.* 1988, 1, 559-565
Loffield R J L F *et al*, *Lancet* 1989i: 1182-1185
B J *et al*, *Gut* 1986, 27, 1132-1137
Goodwin C S, *Lancet* 1988ii 1467-1469
Forman D *et al*, *Int J Cancer* 1990, 46, 608-611
Seppala K M *et al*, *Scand J Gastroenterol* 1992, 27, 973-976
Nomura A *et al*, *N Engl J Med* 1991; 325, 1132-1136
Kosunen T U *et al*, *Lancet* 1992, 339, 893-895

Eden Research Park, Henry Crabb Road, Littleport, Cambridgeshire,
CB6 1SE, UK Tel+ 44(0)1353 862220 Fax+44(0)1353 863330
Email: support@elisa.co.uk Web: www.omegadiagnostics.com
Certified to ISO9001:2008, ISO13485:2003

Genesis Diagnostics Ltd is a subsidiary of Omega Diagnostics Group PLC



H.pylori IgM ELISA Kit

Semi-quantitative / qualitative assay for IgM antibodies to *Helicobacter pylori*

Product Code: GD124

For *in vitro* Diagnostic Use



1. Materials Included in the Kit

- Microplate:** 96 wells in 12 X 8 break-apart strips, pre-coated with partially purified *H.pylori* antigens, with holder in a foil bag with desiccant
- Reagent 1: Sample Diluent** 10mM Tris-buffered saline, pH 7.2 with antimicrobial agent, 46 ml, (blue), read the instructions before use
- IgG absorbent:** Anti-human IgG, 3 x 3.5ml. Read the instructions before use
- Reagent 2: Wash Buffer** 100mM Tris-buffered saline with detergent, pH 7.2, 100 ml, **concentrate** (x10)
- Reagent 3: Conjugate** goat anti-human IgM conjugated to horseradish peroxidase in protein stabilising solution and antimicrobial agent, (green), 12 ml, ready to use
- Reagent 4: TMB Substrate** aqueous solution of TMB and hydrogen peroxide, 12 ml, ready to use
- Reagent 5: Stop Solution** 0.25M sulphuric acid, 12 ml, ready to use
- Standards:** 0 U/ml and 10U/ml, 1ml of 10mM Tris-buffered saline; the 10 U/ml standard also contains human serum IgM antibodies to *H.pylori*, ready to use
- Positive Control:** 1ml of 10mM Tris-buffered saline containing human serum antibodies to *H.pylori*, ready to use
- Negative Control:** 1ml of 10mM Tris-buffered saline containing normal human serum, ready to use
- Instructions for use**

2. Other Equipment Required

Test tubes for dilution • graduated cylinder for preparing wash buffer • precision pipettes and disposable tips to deliver 5µl, 100µl, 1ml • EIA microplate washer or multi-channel pipette or wash bottle • distilled or de-ionised water • absorbent paper • EIA microplate reader with 450nm and optional 620nm reference filter. Alternatively, a suitable, self-validated automated system may be used.

Instrumentation, whether manual or automated, should meet the following criteria: pipettes with better than 3% imprecision with no carry over between pipetting steps; microplate washers should remove 99% of fluid; automated machines should minimise time between washing and adding the next reagent.

3. Intended Use

The *H.pylori* IgM kit is a rapid ELISA method for the detection of IgM antibodies to *Helicobacter pylori*. It is intended for use in clinical laboratories for the evaluation of patients with suspected *H.pylori* infection. The components of the kit are for *in vitro* diagnostic use only.

4. Explanation of the Test

H.pylori infection is a major causative factor in gastritis, gastric and duodenal ulcers and gastric cancer. It is detectable in almost 100% of adult patients with duodenal ulcer and in 90% of patients with gastric ulcer. In developing countries most children become infected by the age of 10 and gastric cancer rates are very high.

The IgM test detects antibodies formed two to four weeks after infection with *H.pylori*. IgG and IgA antibodies require up to two months after infection before becoming detectable. If the immune system fails to eradicate the bacterium, it will begin to colonise the mucosal lining of the upper gastrointestinal tract. In the absence of *H.pylori* IgG, IgM positivity should be correlated with active symptoms of gastric ulcer disease to warrant treatment. If treatment is not chosen in the presence of *H.pylori* IgM antibodies, it is advisable to carry out IgG testing after 2-3 months to confirm infection eradication by the immune system.

The kit uses an antigen preparation that has been partially purified to yield principally high molecular weight *H.pylori* membrane proteins.

5. Principle of the Test

Diluted serum samples are incubated with partially purified *H.pylori* antigens immobilised on microtitre wells. After washing away unbound serum components, rabbit anti-human IgM conjugated to horseradish peroxidase is added to the wells and this binds to surface-bound antibodies in the second incubation. Unbound conjugate is removed by washing, and a solution containing 3,3',5,5'-tetramethylbenzidine (TMB) and enzyme substrate is added to trace specific antibody binding. Addition of Stop Solution terminates the reaction and provides the appropriate pH for colour development. The optical densities of the standards, controls and samples are measured using a microplate reader at 450nm.

6. Safety Precautions

1. All reagents in this kit are for *in vitro* diagnostic use only.
2. Only experienced laboratory personnel should use this test. The test protocol must be followed strictly.
3. CAUTION: the device contains material of human and animal origin and should be handled as a potential transmitter of diseases. All human source material used in the preparation of standards and control for this product have been tested and found negative by ELISA for antibodies to HIV, HbsAg and HCV. No test method, however, can offer complete assurance that infectious agents are absent. Therefore, all reagents containing human material should be handled as if potentially infectious. Operators should wear gloves and protective clothing when handling any patient sera or serum based products.
4. Reagents of this kit contain antimicrobial agents and the Substrate solution contains 3,3',5,5'-tetramethylbenzidine. Avoid contact with the skin and eyes. Rinse immediately with plenty of water if any contact occurs.
5. The Stop Solution contains 0.25M sulphuric acid. Avoid contact with skin and eyes. Rinse immediately with plenty of water if contact occurs.
6. Any liquid that has been brought into contact with potentially infectious material has to be discarded in a container with a disinfectant. Dispose of plates and specimens as clinical waste. Any unused reagents should be flushed away with copious amounts of water. Disposal must be performed in accordance with local legislation.

7. Technical Precautions

1. Strips and solutions should not be used if the foil bag is damaged or liquids have leaked.
2. Allow all reagents and the microplate to reach room temperature before use. Ensure that the microplate foil bag containing any unused strips is well sealed and contains the desiccant to avoid moisture. Store at 2 - 8°C after use.
3. Include the Positive and Negative Control in every test run to monitor for reagent stability and correct assay performance.
4. Strictly observe the indicated incubation times and temperature.
5. When automating, consider excess volumes required for setting up the instrument and dead volume of robot pipette
6. Ensure that no cross-contamination occurs between wells. Keep all pipettes and other equipment used for Conjugate completely separate from the TMB Substrate reagent.
7. When pipetting Conjugate or TMB Substrate, aliquots for the required numbers of wells should be taken to avoid multiple entry of pipette tips into the reagent bottles. Never pour unused reagents back into the original bottles.
8. Do not allow microwells to dry between incubation steps.
9. Strictly follow the described wash procedure. Insufficient washing may cause high background signal.
10. Avoid direct sunlight and exposure to heat sources during all incubation steps.
11. Replace colour-coded caps on their correct vials to avoid cross-contamination
12. It is important to dispense all samples and controls into the wells without delay. Therefore ensure that all samples are ready to dispense.

8. Shelf Life and Storage Conditions

On arrival, store the kit at 2 - 8°C. Once opened the kit is stable for 3 months (or until its expiry date if less than 3 months). Do not use kits beyond their expiry date. Do not freeze any kit component. The diluted Wash Buffer has a shelf life of 3 months if stored in a closed bottle at 2 - 8°C.

9. Specimen Collection and Storage

Serum and plasma samples may be used and should be stored at -20°C for long-term storage. Frozen samples must be mixed well after thawing and prior to testing. Repeated freezing and thawing can affect results. Addition of preservatives to the serum sample may adversely affect the results. Microbially contaminated, heat-treated or specimens containing particulate matter should not be used. Grossly haemolysed, icteric or lipaemic specimens should be avoided.

10. Preparation of Reagents

1. Prepare only sufficient IgG-adsorbent-containing sample diluent for the number of samples to be tested. Add one part IgG adsorbent to 4 parts of Sample Diluent (**Reagent 1**) as shown in the examples below and mix thoroughly. Discard any unused IgG-adsorbent-containing diluent.

Approx. # of Samples	Vol. of Sample Diluent (ml)	Vol. of IgG Adsorbent (ml)
24	10	2.5
48	20	5.0
72	30	7.5
96	40	10.0

2. Dilute the Wash Buffer (**Reagent 2**) 1:9 in distilled water to make sufficient buffer for the assay run e.g. add 50ml wash buffer concentrate to 450ml water.

11. Assay Procedure

1. Dilute patient samples 1:100 in diluted Sample Diluent (5µl serum plus 0.5ml diluent).
2. Assemble the number of strips required for the assay.
3. For semi-quantitative assays, dispense 100 µl of the 0 U/ml standard and 100µl of the 10 U/ml standard, the Negative and Positive Controls and the diluted patient samples into appropriate wells. For qualitative assays, dispense only the 10 U/ml Standard together with controls and samples.
4. Incubate for **30 minutes** at room temperature.
5. After 30 minutes, decant or aspirate the well contents and wash the wells 3 times using automated washing or the manual wash procedure (see below). Careful washing is the key to good results. **Do not allow the wells to dry out.**

Manual Wash Procedure

Empty the wells by inversion. Using a multi-channel pipette or wash bottle, fill the wells with wash buffer. Empty by inversion and blot the wells on absorbent paper. Repeat this wash process 2 more times.

6. Dispense 100µl of Conjugate (**Reagent 3**) into each well. Incubate the wells for **30 minutes** at room temperature.