

## Candida albicans IgA ELISA Kit

Semi-quantitative assay for *Candida albicans* IgA

Product Code: GD019

For *in vitro* Diagnostic Use



### 1. Materials Included in the Kit

- **Microplate:** 96 wells in 12 X 8 break-apart strips, pre-coated with *C.albicans* antigens, with holder in a foil bag with desiccant
- **Reagent 1: Sample Diluent** 10 mM Tris-buffered saline, pH 7.2 with antimicrobial agent, 100ml, (blue), ready to use
- **Reagent 2: Wash Buffer** 100mM Tris-buffered saline with detergent, pH 7.2, 100 ml, **concentrate** (x10)
- **Reagent 3: Conjugate** rabbit anti-human IgA (yellow) conjugated to horseradish peroxidase in protein stabilising solution and antimicrobial agent, 12 ml, ready to use
- **Reagent 4: TMB Substrate** aqueous solution of TMB and hydrogen peroxide, 12 ml, ready to use
- **Reagent 5: Stop Solution** 0.25M sulphuric acid, 12 ml, ready to use
- **Standards:** 0, & 50 U/ml, 1ml of 10mM Tris-buffered saline containing human serum IgA antibodies to *C.albicans*, ready to use
- **Positive control:** 1ml of 10mM Tris-buffered saline containing either human serum IgA antibodies to *C.albicans*, ready to use
- **Instructions for use**

### 11. Assay Procedure Continued...

10. Add 100µl of Stop Solution (**Reagent 5**) to each well. To allow equal reaction times, the Stop Solution should be added to the wells in the same order as the TMB Substrate.
11. Read the optical density (OD) of each well at 450nm in a microplate reader within 10 minutes. A 620nm filter may be used as a reference wavelength for dual wavelength measurements.

### 12. Quality Control

Quality control data is supplied on the lot-specific QC certificate included in the kit.

The positive control is intended to monitor for substantial reagent failure.

Any well positive by spectrophotometer but without visible colour should be cleaned on the underside and re-read. If OD-values below zero are observed, the wavelengths used should be verified, the reader re-blanked to air and the measurements repeated.

### 13. Interpretation of Results

Plot the optical density of the 0 and 50 U/ml standard. Read the unknowns off this line. In our laboratory, we have established normal values to be less than 10 U/ml. Patients with IgA values above 10 U/ml are likely to have had a recent or current *Candida* infection.

### 14. Limitations of the Procedure

1. A positive test does not necessarily indicate infection since antibodies to *Candida* species can be detected in uninfected individuals because of their exposure to commensal yeasts
2. Assays are unable to differentiate antibodies formed during mucosal colonisation from those produced during deep infection. In consequence antibodies are found in many hospitalised patients who have no obvious *Candida* infection.
3. A negative antibody test similarly does not necessarily rule out the possibility of deep-seated candidiasis in immuno-compromised patients who are incapable of mounting an adequate antibody response.
4. Antibody tests are of diagnostic value providing the results are interpreted with caution and in the light of other clinical and laboratory findings. They are most useful when a series of tests is performed at regular intervals over time. A high or rapidly rising antibody titre suggests infection and conversely a reduction in titre indicates recovery.

### 15. Reproducibility

#### Within Assay Precision

CV%: 4 – 8%

#### Between Assay Precision

CV%: 4 – 12%

### 16. Method Summary

- Dilute sera 1:200 with sample diluent (**Reagent 1**)
- Incubate samples in diluent for **30 minutes**
- Dispense standards (as required), the positive control and the diluted samples into the microplate wells
- Incubate for **30 minutes** at room temperature.
- *Wash the wells three times*
- Dispense 100µl of conjugate (**Reagent 3**) into each well
- Incubate at room temperature for **30 minutes**
- *Wash the wells four times*
- Add 100µl of TMB substrate (**Reagent 4**) to each well
- Incubate at room temperature for **10 minutes**
- Add 100µl stop solution (**Reagent 5**) to each well
- Read the optical density at 450nm (single wavelength) or 450/620nm (dual wavelength).

### 17. Further Reading

Faux, J A et al (1992) Comparison of specific IgG antibody levels to the cell wall mannan of *Candida albicans* in normal individuals and in patients with primary antibody deficiency. *J Immunol Methods* 153, 167 - 172

McHugh T M et al (1989) Development of a microsphere based fluorescent immunoassay and its comparison to an enzyme immunoassay for the detection of antibodies to three preparations of *Candida albicans* *J Immunol Methods* 116, 213-219

Burnie J P et al (1985) Outbreak of systemic *Candida albicans* in intensive care unit caused by cross infection. *Brit Med J* 290, 746-748

Crook, W G (1983) The yeast connection: A medical breakthrough. Professional Books, PO Box 3494, Jackson, TN 38301

Robinet, RW Asthma due to *C. albicans* *Univ Michigan, Med Centre Bulletin* 34, 12 - 15

Ray, TL (1980) Fungal infections in the immuno-compromised host. *Med Clin N America* 64, 955 - 968.

Wojdani, A M et al (1986) Measurement of humoral and cellular immunity for the diagnosis of Candidiasis. *Clin Ecol* 4, 210

## 2. Other Equipment Required

Test tubes for dilution • graduated cylinder for preparing wash buffer • precision pipettes and disposable tips to deliver 5µl, 100µl, 1ml • EIA microplate washer or multi-channel pipette or wash bottle • distilled or de-ionised water • absorbent paper • EIA microplate reader with 450nm and optional 620nm reference filter. Alternatively, a suitable, self-validated automated system may be used.

Instrumentation, whether manual or automated, should meet the following criteria: pipettes with better than 3% imprecision with no carry over between pipetting steps; microplate washers should remove 99% of fluid; automated machines should minimise time between washing and adding the next reagent.

## 3. Intended Use

The *Candida albicans* IgA kit is a rapid ELISA method for the detection of IgA antibodies to *Candida albicans*. It is intended for the diagnosis of *Candida albicans* infection. The components of the kits are for *in vitro* diagnostic use only.

## 4. Explanation of the Test

A number of fungal pathogens can spread systemically from the intestinal lumen to the visceral organs. Although hundreds of *Candida* species are known, *C. albicans* and *C. tropicalis* cause over 80% of *Candida* infections in man. *Candida* probably enters the newborn in the first days of life and is a normal inhabitant of the gut. Systemic candidiasis is a fungal infection of the deep organs resulting from the overgrowth and spread of *Candida*. It is a significant cause of death in immuno-compromised patients or those undergoing prolonged antibiotic therapy. *Candida* infection is not routinely tested for in blood donors and may also be transmitted via blood transfusions. Elevated *Candida* IgG levels are also frequently encountered in elderly patients (>60 years).

Recent infection with systemic candidiasis is characterised by elevated IgM and IgG titres. *Candida* IgA antibodies are associated with mucosal membrane infections. We recommend that patients with suspected candidiasis are screened for all three antibody classes.

## 5. Principle of the Test

Diluted serum samples are incubated with *C. albicans* antigens immobilised on microtitre wells. After washing away unbound serum components, rabbit anti-human IgA conjugated to horseradish peroxidase is added to the wells, and this binds to surface-bound antibodies in the second incubation. Unbound conjugate is removed by washing, and a solution containing 3,3',5,5'-tetramethylbenzidine (TMB) and enzyme substrate is added to trace specific antibody binding. Addition of stop solution terminates the reaction and provides the appropriate pH for colour development. The optical densities of the standards, positive control and samples are measured using a microplate reader at 450nm.

## 6. Safety Precautions

1. All reagents in this kit are for *in vitro* diagnostic use only.
2. Only experienced laboratory personnel should use this test. The test protocol must be followed strictly.
3. CAUTION: the device contains material of human and animal origin and should be handled as a potential transmitter of diseases. All human source material used in the preparation of standards and control for this product have been tested and found negative by ELISA for antibodies to HIV, HbsAg and HCV. No test method, however, can offer complete assurance that infectious agents are absent. Therefore, all reagents containing human material should be handled as if potentially infectious. Operators should wear gloves and protective clothing when handling any patient sera or serum based products.
4. Reagents of this kit contain antimicrobial agents and the Substrate solution contains 3,3',5,5'-tetramethylbenzidine. Avoid contact with the skin and eyes. Rinse immediately with plenty of water if any contact occurs.
5. The Stop Solution contains 0.25M sulphuric acid. Avoid contact with skin and eyes. Rinse immediately with plenty of water if contact occurs.
6. Any liquid that has been brought into contact with potentially infectious material has to be discarded in a container with a disinfectant. Dispose of plates and specimens as clinical waste. Any unused reagents should be flushed away with copious amounts of water. Disposal must be performed in accordance with local legislation.

## 7. Technical Precautions

1. Strips and solutions should not be used if the foil bag is damaged or liquids have leaked.
2. Allow all reagents and the microplate to reach room temperature before use. Ensure that the microplate foil bag containing any unused strips is well sealed and contains the desiccant to avoid moisture. Store at 2 – 8°C after use.
3. When automating, consider excess volumes required for setting up the instrument and dead volume of robot pipette
4. Include the positive control in every test run to monitor for reagent stability and correct assay performance.
5. Strictly observe the indicated incubation times and temperature.
6. Ensure that no cross-contamination occurs between wells. Keep all pipettes and other equipment used for enzyme conjugate completely separate from the substrate reagent.
7. When pipetting Conjugate or Substrate, aliquots for the required numbers of wells should be taken to avoid multiple entry of pipette tips into the reagent bottles. Never pour unused reagents back into the original bottles.
8. Do not allow microwells to dry between incubation steps.
9. Strictly follow the described wash procedure. Insufficient washing may cause high background signal.
10. Avoid direct sunlight and exposure to heat sources during all incubation steps.
11. Replace colour-coded caps on their correct vials to avoid cross-contamination
12. It is important to dispense all samples and the positive control into the wells without delay. Therefore ensure that all samples are ready to dispense.

## 8. Shelf Life and Storage Conditions

On arrival, store the kit at 2 - 8°C. Once opened the kit is stable for 3 months (or until its expiry date if less than 3 months). Do not use kits beyond their expiry date. Do not freeze any kit component. The diluted Wash Buffer has a shelf life of 3 months if stored in a closed bottle at 2 - 8°C.

## 9. Specimen Collection and Storage

Serum and plasma samples may be used and should be stored at -20°C for long-term storage. Frozen samples must be mixed well after thawing and prior to testing. Repeated freezing and thawing can affect results. Addition of preservatives to the serum sample may adversely affect the results. Microbially contaminated, heat-treated or specimens containing particulate matter should not be used. Grossly haemolysed, icteric or lipaemic specimens should be avoided.

## 10. Preparation of Reagents

Dilute the Wash Buffer (**Reagent 2**) 1:9 in distilled water to make sufficient buffer for the assay run e.g. add 50ml wash buffer concentrate to 450ml water.

## 11. Assay Procedure

1. Dilute patient samples 1:200 in sample diluent (e.g. 5µl serum plus 1ml diluent).
  2. Allow the diluted samples to stand for **30 minutes**. The sample diluent contains an immunoabsorbent that removes non-candida antibodies
  3. Assemble the number of strips required for the assay.
  4. Dispense 100 µl of each standard, and positive control and the diluted patient samples into appropriate wells.
  5. Incubate for **30 minutes** at room temperature.
  6. After 30 minutes, decant or aspirate the well contents and wash the wells 3 times using automated washing or the manual wash procedure (see below). Careful washing is the key to good results. **Do not allow the wells to dry out.**
- Manual Wash Procedure  
Empty the wells by inversion. Using a multi-channel pipette or wash bottle, fill the wells with wash buffer. Empty by inversion and blot the wells on absorbent paper. Repeat this wash process 2 more times.
7. Dispense 100µl of Conjugate (**Reagent 3**) into each well. Incubate the wells for **30 minutes** at room temperature.
  8. After 30 minutes, discard the well contents and carefully wash the wells 4 times with wash buffer. Ensure that the wells are empty but do not allow to dry out.
  9. Using a repeating dispenser, rapidly dispense 100µl of TMB Substrate (**Reagent 4**) into each well. Incubate the plate for **10 minutes**.