

Autonomy Control Sera

CODE GDA08 THYROGLOBULIN AND THYROID PEROXIDASE
ANTIBODIES

For *in vitro* use only

250908

Applications: Un-assayed controls

Positive Control GDA-08TGP

Description: Human serum containing antibodies to thyroglobulin (TG) with added stabilisers and sodium azide as preservative (1 vial containing 0.5ml)

Positive Control GDA-08TPOP

Description: Human serum containing antibodies to thyroid peroxidase (TPO) with added stabilisers and sodium azide as preservative (1 vial containing 0.5ml)

Donor Pathology: Autoimmune thyroiditis

Negative Control GDA-0N

Description: Normal human serum with added stabilisers and sodium azide as preservative (1 vial containing 0.5ml)

Storage

Store product at 2-8°C; once opened the product has a shelf- life of 120 days

Safety Precautions

The human sera contained in this product have been tested and found negative for HCV, HbSag and HIV 1&2 by ELISA methods. Since no one test is 100% accurate, the product should be handled as if potentially infectious. Operators should wear gloves and protective clothing when handling these products. This product contains 0.09% sodium azide as preservative and is classified as non-hazardous according to EC Directives.

Genesis Diagnostics Ltd., Eden Research Park, Henry Crabb Road, Littleport
Cambridgeshire, CB6 1SE, United Kingdom.
Tel: +44 1353 862220 Fax: +44 1353 863330
Web: www.elisa.co.uk email: genesis@elisa.co.uk

A subsidiary of Omega Diagnostics Group plc